

Aberystwyth University

Computer-aided diagnosis

Rampun, Yambu Andrik; Chen, Zhili; Malcolm, Paul; Tiddeman, Bernie; Zwiggelaar, Reyer

Published in:

International Journal for Numerical Methods in Biomedical Engineering

DOI:

[10.1002/cnm.2745](https://doi.org/10.1002/cnm.2745)

Publication date:

2016

Citation for published version (APA):

Rampun, Y. A., Chen, Z., Malcolm, P., Tiddeman, B., & Zwiggelaar, R. (2016). Computer-aided diagnosis: Detection and localization of prostate cancer within the peripheral zone. *International Journal for Numerical Methods in Biomedical Engineering*, 32(5), [e02745]. <https://doi.org/10.1002/cnm.2745>

General rights

Copyright and moral rights for the publications made accessible in the Aberystwyth Research Portal (the Institutional Repository) are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Aberystwyth Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Aberystwyth Research Portal

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

tel: +44 1970 62 2400
email: is@aber.ac.uk

Computer Aided Diagnosis: Detection and Localisation of Prostate Cancer within the Peripheral Zone

Andrik Rampun*, Zhili Chen**, Paul Malcolm***, Bernie Tiddeman*,
Reyer Zwiggelaar*

**Department of Computer Science, Aberystwyth University, Aberystwyth SY23 3DB, UK*

***Information and Control Engineering Faculty, Shenyang Jianzhu University, Liaoning 110168, China*

****Department of Radiology, Norfolk Norwich University Hospital, Norwich NR4 7UY, UK*

SUMMARY

We propose a methodology for prostate cancer detection and localisation within the peripheral zone based on combining multiple segmentation techniques. We extract four image features using Gaussian and median filters. Subsequently, we use each image feature separately to generate binary segmentations. Finally, we take the intersection of all four binary segmentations, incorporating a model of the peripheral zone, and perform erosion to remove small false positive regions. The initial evaluation of this method is based on 275 MRI images from 37 patients and 86% of the slices were classified correctly with 87% and 86% sensitivity and specificity achieved, respectively. This paper makes two contributions: firstly, a novel Computer Aided Diagnosis approach which is based on combining multiple segmentation techniques using only a small number of simple image features. Secondly, the development of the proposed method and its application in prostate cancer detection and localisation using a single MRI modality with the results comparable to the state-of-the-art multi-modality and advanced computer vision methods in the literature. Copyright © 2015 John Wiley & Sons, Ltd.

KEY WORDS: Prostate cancer detection, MRI, Prostate cancer localisation

1. INTRODUCTION

Prostate cancer is the most commonly diagnosed cancer among men and remains the second leading cause of cancer death in men globally. In 2013, there were approximately 240,000 and 40,000 cases reported in the United States and United Kingdom, respectively, and is estimated to reach 1.7 million cases globally by 2030 [1, 2]. In the last decade, prostate cancer screening has been receiving more attention because it can assist in detecting cancer at an early stage before there are any externally detectable symptoms. Statistically, nine out of ten men survive for at least five years if the cancer is diagnosed at the earliest stage [3]. However, early detection of prostate cancer remains a source of uncertainty and controversy [3]. Clinical diagnostic tools such as prostate-specific antigen (PSA) level, digital rectal examination (DRE), transrectal ultrasound (TRUS) and biopsy tests are very popular and globally used despite their inconsistency in producing accurate results [4]. According to Schroder et al. [5], although the use of PSA reduces the rate of death by 20%, the benefit was associated with a high risk of overdiagnosis and overtreatment. It should be noted that the PSA test is not able to predict the aggressiveness of cancer. As a result, slow-growing and non-aggressive prostate cancer is frequently diagnosed in older patients [6]. In terms of TRUS-guided biopsy,

*Correspondence to: A. Rampun, R. Zwiggelaar. Department of Computer Science, Aberystwyth University, Aberystwyth SY23 3DB, UK

†Email: yar@aber.ac.uk, rrz@aber.ac.uk

many reports have shown that systematic biopsies do not detect all clinically significant cancers: [6] showed in a large study that nearly a quarter (23%) of detectable cancers were missed [7]. Finally, according to [8] DRE is less effective than the PSA blood test in finding prostate cancer, but it can sometimes find cancers in men with normal PSA levels. One common problem with DRE is if a tumor is located away from the rectal wall, the physician will not be able to palpate it. A recent study concluded that an abnormal DRE had 0.44 sensitivity and 0.68 specificity [9]. On the other hand, although TRUS enables the accurate determination of prostate size and depicts zonal anatomy, its ability to detect cancer tissue is limited with sensitivity and specificity varying between 40-50% [10].

Computer Aided Diagnosis (CAD) of prostate magnetic resonance imaging (MRI) has the potential to improve the accuracy, sensitivity and specificity of clinical diagnostics. According to [11], it could achieve a sensitivity ranging from 61%-81% (average: 71%), specificity 91%-96% (average: 93.5%) and accuracy 84%-92% (average: 88%) while being a non-invasive technique. Moreover, a study conducted in [6] has shown that combining anatomical, functional and metabolic MRI information could achieve on average 83% (75%-92%) sensitivity. Unfortunately, the assessment of prostate MRI requires a high level of expertise and suffers from observer variability [6]. CAD systems can be of benefit to improve the diagnostic accuracy of radiologists, reduce variability and speed up the reading time [6]. The initial goal of CAD is to automatically delineate malignant regions, leading to a reduction of search and interpretation errors, as well as a reduction of the variation between and within observers [6]. CAD has been successfully implemented in different fields of medical imaging such as mammography [12], CT chest [13], CT colonography [14] and brain imaging [15]. Figure 1 shows an example MRI image with the ground truth of the prostate gland, central zone and tumor represented in yellow, green and red, respectively (left image), while the right image shows a simpler schematic overview of the prostate derived from the prostate anatomy proposed in [72] with central zone (CZ), peripheral zone (PZ), and tumor (T). Note that the transitional zone (TZ) is located within the CZ but no definable boundary between these regions is expected on MRI. The ultimate goal of this research is to develop a CAD tool




Figure 1. Prostate MRI image (left image) with its ground truth delineated by an expert radiologist and a schematic (right image) overview of the prostate containing a tumor.

for prostate cancer detection and localisation within the PZ mainly because a) about 80% of the prostate cancers appear in the PZ [10, 16, 17, 18] and b) in general, prostate cancer that arises in the peripheral zone is more aggressive than that which arises in the transitional or central zones [74]. Therefore, in this paper we propose a new method for detecting prostate abnormality within the PZ using four different image features (see Figure 3) extracted using Gaussian and median filters. The main goal of this method is to identify malignant regions (and hence localise them) within the PZ by taking the overlapping binary segmentation from each image feature. This means pixels (or tissues) which are classified in the same malignant cluster in all image features are considered to have the highest probability of being malignant. However, if a pixel is classified as belonging to a benign or normal tissue cluster in one of the image features we considered it to be a benign or normal tissue (this will be explained in more detail in section 4).

The novelty of this method resides in an approach which combines simple features (this is similar to a forest of weak classifiers which together provide strong results), for the first time and applied to prostate T2-W using one modality. To our knowledge, no existing methods in the literature have used the technique of finding cancer regions by taking the overlapping binary segmentation extracted from a small number of image features.

2. RELATED RESEARCH AREA

CAD is a valuable tool and becoming increasingly important in assisting and automating specific clinician's tasks such as detection, localisation, the study of anatomical structure, treatment planning and computer-assisted surgery [75]. We are focusing on developing a CAD tool for the detection and localisation of abnormal region within the PZ in T2-MRI imaging. CAD can be applied in many different medical imaging applications (such as brain, breast, chest and prostate) using different segmentation techniques. According to [93] the most commonly used segmentation techniques in CAD systems could be categorised into six groups:

1. Contour and shape based (e.g. Active Contour, Level Set, Graph Searching, Atlas-based, Deformable models, etc).
2. Machine learning based (e.g. Support Vector Machine (SVM), k-Nearest Neighbors (k-NN), Fuzzy C-Mean (FCM), K-means, etc)
3. Region based (e.g. Thresholding, Edge-based, watershed, split and merge, etc)
4. Statistical based (e.g. Markov Random Field (MRF), Gaussian Mixture Model (GMM), etc)
5. Multiresolution based analysis (e.g. Discrete Wavelet Transform (DWT), etc)
6. Hybrid and soft computing methods (e.g. Level Set + Artificial neural network (ANN), Fuzzy C-means + DWT, etc)

The most popular segmentation techniques in biomedical imaging fall under the supervised and unsupervised machine learning based techniques, and contour and shape based methods. In this section we will briefly discuss these techniques and their applications in biomedical imaging. For the other techniques and their applications in medical imaging we refer to [11, 12, 13, 14, 15, 96].

The level set technique has been applied to several human organs (e.g. brain, cardiac, prostate, breast, etc). Dubey et al. [77] proposed a semi-automatic segmentation method of MRI brain tumors. Firstly, the method generated a tumor probability map by classifying each voxel into the tumor or background class using intensity-based fuzzy c-means. Subsequently, the tumor probability map was used to locally guide the propagation direction of the level set. Tsai et al. [78] developed a shape-based approach using level sets and demonstrated their method by applying it to the segmentation of cardiac and prostate MRI. The proposed method derived a parametric model for an implicit representation of the segmenting curve by applying principal component analysis to a collection of signed distance representations of the training data. The parameters of this representation were then manipulated to minimize an objective function for segmentation. Liu et al. [79] proposed a method for mass segmentation in mammograms using a level set to improve the initial segmentation performed using a watershed algorithm. On the other hand, Shi et al. [80] used k -means clustering followed by a morphological opening operation for initial mass segmentation. For the level set segmentation a linear discriminant analysis (LDA) classifier with stepwise feature selection was used to merge the extracted features into a classification score.

Recently, Yeo et al. [81], proposed a level-set segmentation method using active contour modelling applied to synthetic and real images (e.g. brain and knee MRI and carotid CT image). The proposed method consisted of an image attraction force, which was used to propagate contours toward object boundaries, and a global shape force, which deforms the model according to the shape distribution learned from a training set. On the other hand, Sachdeva et al. [82] proposed a method which used intensity and texture information (extracted from Gray-Level Co-occurrence Matrices) present within the active contour to overcome weak or diffused edges in an image. In [83], a novel automatic approach to identify brain structures in magnetic resonance imaging (MRI) is presented for volumetric measurements. This approach combines the active contour model with a support vector machine (SVM) classifier. The SVM features are selected according to the structure of brain tissues: gray matter (GM), white matter (WM), and cerebrospinal fluid (CSF). Jones et al. [95] developed an interactive segmentation method by combining both region selection and user point selection. Evaluation results showed on average more than 98% accuracy based on 248 intravascular ultrasound (IVUS) images.

Graph searching techniques have been studied for the segmentation of biomedical images such as brain, knee and glottis [84, 85, 86]. Padoia and Binaghi [84] proposed a fully automatic 2D

brain segmentation using graph searching technique which consisted of border detection based on two-dimensional graph searching principles and radial contour detection. In the border detection phase, polar conversion is performed first, followed by skull and brain boundaries detection. Li et al. [85] extended the optimal graph-searching techniques to 3D and higher dimensions using a polynomial-time algorithm for surface segmentation in volumetric images. The method is efficient and robustly tested on MR arterial walls and IVUS image. Finally, a study in [86] uses the output of SVM to drive a graph-cuts segmentation, which was initially trained as a local Golgi detector based on rotationally invariant features. It should be noted that in some of the described segmentation approaches difference images have been used but in general these rely on multi-modality data or on some temporal sequences, neither of which have been used in the proposed approach. In the developed approach we rely on a single modality and the variation in appearance between normal and abnormal tissues within the peripheral zone.

Despite the relevant achievements obtained, the main limitations of contour and shape based methods for our research are

1. Most of them require a significant amount of user interaction for initial region selection. In our case, we want to eliminate (or minimise) user interaction in finding cancerous regions.
2. They work well only if the boundary of the object is well defined within the image: such methods work well for prostate gland detection [94]. In many prostate T2-W MRI, cancerous regions are vague both in terms of appearance and shapes, and as such many training samples would be needed.
3. The results of many of these methods are highly dependent on the initialisation.

Therefore, we applied an unsupervised machine learning based method: FCM clustering as proposed by Chen and Zwigelaar [66]. This method incorporates local spatial and intensity information based on an adaptive local window filter whose weighting coefficients differentiate the neighbouring pixels within the local window. The method is less sensitive in dealing with different types of noise and intensity inhomogeneities. Several noise reduction techniques such as median and Gaussian filters were used to reduce image noise and erosion is used to reduce false positives.

3. MODELLING THE PERIPHERAL ZONE

Pathologically, about 80% of prostate cancers arise in the PZ and the rest are within the CZ [16]. Since the percentage occurrence of cancers in the PZ is high and as these tend to be more aggressive, we aim to detect prostate abnormality within this region. We did not perform prostate segmentation because all prostates were already delineated by an expert. It should be noted that Zhu et al. [94] developed a method to detect the prostate capsule. Based on the schematic overview shown in the right image of Figure 1 (also proposed in [72]), we defined our 2D prostate model based on Figure 2.




Figure 2. Prostate gland (black) and the defined PZ below $y = ax^2 + bx + c$ (green) which goes through v_1, v_2 and v_3 .

The generic prostate's PZ model in this paper is mainly inspired from similar models proposed by Makni et al. [73] and Liu et al. [91] which used catenary and polynomial curves, respectively.

From a radiological point of view, the prostate is mainly divided into two regions in MR (the PZ and CZ). Therefore, according to Makni et al. [73], when segmenting these regions, expert radiologists tend to follow the rule of 'imagining' outlines due to contrast or strong artifacts. The process of distinguishing these regions are heavily relying on a *priori* knowledge of their most likely locations [73]. Indeed, a more accurate way (probably more complex and time consuming) could be achieved by segmenting the PZ within the prostate gland. However, these approaches are complicated and require high accuracy in distinguishing tissues in the PZ and CZ. In cases where there is no clear boundary between the PZ and CZ, most segmentation based methods suffer from over-segmentation (hence, could lead to many false positives). In contrast, defining quadratic curves is simple and fast.

In our CAD system, we used the quadratic equation $y = ax^2 + bx + c$ based on three crucial coordinate points of the prostate which are v_1, v_2 and v_3 , which are determined by the outmost x and y coordinates of the prostate boundary: $x_{min}, x_{max}, y_{min}, y_{max}$ (see Figure 2). For example, x_{min} and y_{max} can be determined by taking the minimum x and maximum y coordinates along the prostate boundary. Moreover, the x coordinates of v_1 and v_3 are captured from x_{min} and x_{max} and their y coordinate is determined by taking the y coordinate between y_{min} and y_{max} . On the other hand, the x coordinate of v_2 is taken from the x coordinate x_{min} and x_{max} and its y coordinate is determined by taking $\frac{3}{4}$ (0.75) of the distance from y_{min} to y_{max} . The coefficient ($\epsilon = 0.75$) is selected as it gives balanced results in terms of accuracy, sensitivity and specificity (see Figure 13). Mathematically, these can be represented in equations (1), (2), (3) and (4).

$$C_p = ((x_{min} + x_{max})/2, (y_{min} + y_{max})/2) \quad (1)$$

$$v_1 = (x_{min}, (y_{min} + y_{max})/2) \quad (2)$$

$$v_2 = ((x_{min} + x_{max})/2, y_{min} + ((y_{max} - y_{min}) \times \epsilon)) \quad (3)$$

$$v_3 = (x_{max}, (y_{min} + y_{max})/2) \quad (4)$$

Once the coordinates of v_1, v_2 and v_3 are defined, we can determine the values of a, b and c (therefore a final quadratic equation is defined). Finally, by taking every x coordinate from x_{min} to x_{max} into a quadratic equation we are able to determine the y coordinate which will define the PZ's boundary (the main goal is to analyse the region under the green line in Figure 2). The approximation model is able to capture most of the PZ area, easy to implement and computationally efficient.

4. METHODOLOGY

Figure 3 shows the overview of the proposed methodology. First, we perform Gaussian and median filtering on the original image to obtain G_1 and M_1 . We extract a probability image from G_1 and M_1 using greyscale frequency before we obtain the third feature (F_1) which is the magnitude of G_1 and M_1 . This means, each element in F_1 is the sum vector of each component from G_1 and M_1 . On the other hand, the fourth feature (F_2) is the vector magnitude of probability images from G_2 and M_2 . Subsequently, we use each feature separately to generate binary segmentations taking the feature space and intensity values into account. We perform erosion on each of the segmentations to remove small false positive regions. Finally, we take the intersection of all four binary segmentations, taking a model of the peripheral zone (see Section 3) into account.

4.1. Preprocessing

Since MRI images often suffer from different types of noise, it is necessary to apply different types of denoising methods before doing any further processing. From a clinical point of view, this is an appropriate step to enhance characteristics of a important region of interest (such as textures and boundaries). Moreover, it does not deform the anatomical locations of tissue regions because this step deals only with noise without affecting the spatial information.

Hendrick [59] reported that one type of noise in MRI images is Gaussian noise. Therefore, Gaussian smoothing is selected, which is also effective in reducing noise that is problematic for




Figure 3. Overview of the proposed methodology. Intensity values are represented by greyscale colours, with the darkest representing the lowest intensity.

image analysis algorithms. For example, image segmentation is often affected by the presence of too many local minima/maxima and inflection points in the data [19]. Studies performed by Barentsz et al. [89] and Viswanath et al. [90] suggested that most cancers shows textural distortions in T2-W images. Litjens et al. [92] captured these characteristics using Gaussian filters. Many previous studies have applied Gaussian filters in denoising MRI images and mammograms [59, 60, 61]. We are aware that there are more sophisticated denoising methods such as those based on Fourier analysis [87] or anisotropic filtering [88] which could be explored in future work. The 2D Gaussian function is defined as

$$g(s, t) = \frac{1}{2\pi\sigma^2} \cdot e^{-\frac{s^2+t^2}{2\sigma^2}} \quad (5)$$

where s is the distance from the origin in the horizontal direction, t is the distance from the origin in the vertical direction, and σ is the standard deviation of the Gaussian distribution. In the proposed method we used the following parameters: the kernel size (ks) is 15×15 and the standard deviation (σ) is 3.0. See subsection 4.2 for the selection of the Gaussian parameters (σ and ks) and their resulting variability can be seen in Figure 12.

On the other hand, we used median filtering to preserve the regional boundaries (e.g. tumor regions). It is claimed that using median filtering is much better at preserving sharp edges [30] and in our case we want to preserve the information-bearing structures such as tumor boundaries [10]. The median filter works by replacing the pixel value with the median value in the neighborhood of that pixel. We used a sliding window of 5×5 pixels. Other sizes are possible (such as 3×3 , 7×7 and 9×9) give similar results.

We calculated the probability images using equation (6) and calculate the vector sum using equation (7). Probability images are commonly used to model the expected appearance of an object (e.g. tumor region) in a given reference space. Many studies suggested that prostate cancer appears darker within the PZ and is similar to the tissues outside the prostate gland. By computing a

probability image we are able to quantify the likelihood of every pixel/voxel belonging to specific tissues (e.g. tumor region). This also means each pixel/voxel can be represented by a value of the likelihood of being malignant. The conversion to probability images acts as a normalisation across the feature images and as such these can all be treated in the same framework. We calculate the probability image for each of G_1 and M_1 . This means, for an $f(i, j)$ image, the probability value for the k^{th} grey level is calculated using:

$$P(i, j) = \frac{\#(f(i, j) = k)}{M \times N} \quad (6)$$

where $\#(f(i, j) = k)$ is the number of pixels at the k^{th} intensity level in a $M \times N$ image, and as such each element in P is the probability value for a particular intensity level. Others have exploited probability images for segmentation [62, 63]. To calculate features F_1 and F_2 , we find the sum vector for every corresponding element in G_1 and M_1 using equation (7). According to [65] by taking the square root of the sum of squares between two corresponding signals produces a good image with little noise and continuous edge marking while another study performed in [64] suggested that equation (7) can improve the signal to noise ratio (SNR). In this case, corresponding signals are the corresponding pixels in G_1 and M_1 .

$$I_n(i, j) = \sqrt{G_n^2(i, j) + M_n^2(i, j)}, n = 1 \text{ or } 2 \quad (7)$$

In total, four features are extracted, namely the Gaussian feature (G_1), the median feature (M_1), the magnitude Gaussian and median features (F_1) and the magnitude of the probability images of Gaussian and median features (F_2). Before image segmentation is performed, we applied noise reduction to F_1 and F_2 to minimise the noise retained/created after being processed using equations (6) and (7). In the proposed method we applied a robust noise reduction method developed by Garcia [20] which is robust in dealing with weighted, missing, and outlying values by using an iterative procedure (which is the case in G_2 and M_2). Figure 4 shows examples of all extracted features G_1 , F_1 , F_2 and M_1 . We can see that the malignant region appears brighter in F_2 and darker in the other features.




Figure 4. Example extracted features, from left to right: G_1 , F_1 , F_2 and M_1 . The outline prostate is defined by the yellow line and the malignant region (in the PZ) is indicated by the red arrow.

4.2. Gaussian parameters

The selection of the parameters for the Gaussian smoothing function is based on the studies in [23, 24, 25, 26, 27], which indicated that, the standard deviation (σ) and kernel size (ks), are linked. According to the experiments with different Gaussian convolution algorithms conducted in [23], the authors showed that the amount of error (the smaller the error the closer the denoised image is in comparison to the original image) did not change significantly after $\sigma \geq 2$. This means, for many Gaussian algorithms the error is much higher (less accurate) when $\sigma < 2$. On the other hand, for the selection of kernel size (ks) several authors [24, 25, 26, 27] suggested that in general, filter size should be $\lceil 3\sigma \rceil$ to $\lceil 5\sigma \rceil$ and odd [27, 28]. For instance, if $\sigma = 2.5$, the recommended minimum kernel size is 9 ($3 \times 2.5 = \lceil 7.5 \rceil = 8$, since it should be an odd number according to [27, 28], the nearest odd value is 9). Similarly, selecting $\sigma = 1$ would suggest the smallest kernel size of 3×3 . Although there are no quantitative experimental results for optimal Gaussian parameters on medical images such as MRI or ultrasound, their results indicate a general guideline for selecting Gaussian parameters. In the proposed method we used several σ values together with several kernel sizes

and chose the ones that give the highest accuracy, sensitivity and specificity (see Figure 12). The selection of parameters is not the major focus of this study but the development of a novel method of prostate cancer detection and localisation within the PZ is.

4.3. Clustering

In the proposed method, image segmentation is performed using a Fuzzy C-Means (FCM) algorithm as it has been widely applied in a variety of medical image segmentation applications [29, 66]. However, one common problem with FCM is its ability in handling different types of noise and intensity inhomogeneities taking local spatial and intensity information into account. The FCM algorithm assumes that every pixel can belong to multiple classes with varying degrees of membership. The algorithm works by assigning membership to each data point corresponding to each cluster center on the basis of distance between the cluster and the data point. The closer the data point to the cluster center the higher its membership value for that cluster. Let $Y = (y_1, y_2, \dots, y_R)$ denote an image with R pixels to be partitioned into d clusters. FCM iteratively minimises the objective function defined as

$$J_{fcm} = \sum_{p=1}^d \sum_{q=1}^R u_{pq}^m \|y_q - v_p\|^2 \quad (8)$$

with the following constraints: $\sum_{p=1}^d u_{pq} = 1$ for $\forall q$ and $0 < \sum_{q=1}^R u_{pq} < R$ for $\forall p$, where u_{pq} represents the membership of pixel y_q to the p^{th} cluster, y_q represents the feature data of the q^{th} pixel, and v_p is the prototype value of the p^{th} cluster centre. The parameter m (equal to 2 in this study) is a weighting exponent on each fuzzy membership that controls the fuzziness of the resulting partition. We segment every image feature into four different classes. We selected four classes based on the number of tissue categories in the prostate: normal (non-neoplastic) prostatic tissue, benign prostatic hyperplasia, high-grade prostatic intraepithelial neoplasia, and prostatic adenocarcinoma [52]. The first two categories are benign tissues, the third one is a risk factor for malignancy (we included this into one of the malignant classes to reduce false negatives) and the last one is malignant. Most cancer regions in the PZ tend to have a dark appearance [16, 31]. Moreover, several studies suggested that prostate cancer tissue tends to appear darker on a T2-weighted MRI image [32, 33, 34]. In fact, radiologists also tend to use darker regions to identify abnormality within the PZ [35]. Since most malignant regions contain lower intensities, cancerous regions could be detected within the prostate by taking the segmented regions that correspond to the first two lowest intensity fuzzy c-means clusters (indicated by the superscript 'low' in Figure 5) in G_1 , F_1 and M_1 . However, since malignant regions in F_2 are represented by higher average intensity values, we take segmented regions which correspond to the two highest intensity fuzzy c-means clusters. This process can be represented using the following equation

$$O = G_1^{low} \cap F_1^{low} \cap F_2^{high} \cap M_1^{low} \quad (9)$$

where 'low' and 'high' are low and high intensity represented in the segmented regions within the prostate and O represents the overlapping region from all four binary segmentations. Figure 5 shows an example of this process.

After selecting the regions of interest (segmented areas which are under the approximate PZ's boundary (green line in Figure 2)), we combine all binary segmentations and find its overlapping region as shown in Figure 5. Finally, we perform erosion to remove noisy pixels which will be explained in the next subsection. Note that segmented areas above the green line were ignored in this study because we are only interested in detection within the PZ.

4.4. Post processing

By performing erosion on the binary segmentation, we can reduce the number of false positives. The number of pixels removed from the objects in an image depends on the size and shape of the structuring element used. In the proposed method we used a 'disk' shaped structuring element with




Figure 5. After performing FCM clustering on every feature, we only take segmented regions which correspond to the first two lowest intensity FCM in G_1 , F_1 and M_1 , and segmented regions which correspond to the two highest intensity clusters in F_2 . Note that only segmented areas which are under the green line in Figure 2 (the estimation of the PZ's boundary) will be taken into account. Segmented areas above the green line were removed.

size either 1 or 2. The size selection of the structuring element depends on the size of the segmented region within the peripheral zone. If the size of the segmented region within the peripheral zone covers $\geq 20\%$ of the size of the peripheral zone, the size of the structuring element is 2, otherwise 1. This ensures that every segmented region is not over eroded or under eroded during the process.

5. DATABASE DESCRIPTION

Data from 37 patients (range: 40-74 years) with biopsy-proven prostate cancer were included in this study. All patients underwent T2-W MR imaging at the Department of Radiology at the Norfolk and Norwich University Hospital, Norwich, UK. MR acquisitions were performed prior to radical prostatectomy. All patients gave their written consent to participate in this study which was approved by the institutional review board. All images were obtained on a 1.5 Tesla magnet (Sigma, GE Medical Systems, Milwaukee, USA) using a phased array pelvic coil, with a 24×24 cm field of view, 512×512 matrix, 3mm slice thickness, and 0.5mm inter-slice gap. Each patient has 5 to 12 slices. However, since our current study is focusing only within the PZ, slices with no visible PZ (the whole prostate gland is covered by the CZ) were excluded in this study (e.g. see Figure 6). All images were manually annotated by an expert radiologist (and further validated/confirmed by two independent radiologists) with more than 10 years experience in diagnosing prostate cancer in MRI. In total our database contains 275 slices (135 malignant and 140 normal slices). Each slice contains the annotations of prostate gland, central zone and cancerous regions (if present).




Figure 6. The whole of the prostate gland is fully covered by the CZ. All cases like this were excluded in our study because our current focus is within the PZ.

6. EXPERIMENTAL RESULTS

Data was analysed and classified as to whether the prostate contains cancer. The detection of cancer occurs when there are any retained segmented regions ($G_1^{low} \cap F_1^{low} \cap F_2^{high} \cap M_1^{low}$) within the peripheral zone. Subsequently, we compared the result with the ground truth whether the prostate contains cancer regions or not. We use several quantitative measures to evaluate the results such as sensitivity, specificity and accuracy. Each of these metrics can be calculated using the following equations

$$Sensitivity(Sen) = \frac{TP}{TP + FN} \quad (10)$$

$$Specificity(Spe) = \frac{TN}{TN + FP} \quad (11)$$

$$Accuracy(Acc) = \frac{TP + TN}{TN + TP + FP + FN} \quad (12)$$

where TP and FP denote the number of true positives and false positives, respectively. Similarly, TN and FN indicate the numbers of true negatives and false negatives. Accuracy means the number of correct classified slices (or pixels in voxel based classification) out of the total number of slices. Sensitivity measures the proportion of actual positives which are correctly identified (in this case the percentage of malignant slices which are correctly identified) whereas specificity measures the proportion of actual negatives which are correctly identified (in this study the percentage of normal slices which are correctly identified). The proposed method achieved 86% accuracy (237 samples are classified correctly) and 38 samples data are misclassified with 7% (20 samples) false negative and 6% (18 samples) false positive results. In addition, the method produced 87% sensitivity and 86% specificity. On the other hand, in terms of voxel based classification (only within the PZ) we achieved 0.86 ± 0.06 , 0.83 ± 0.05 and 0.96 ± 0.07 accuracy, sensitivity and specificity. Erosion with flexible size of structuring element and regions intersection ($G_1^{low} \cap F_1^{low} \cap F_2^{high} \cap M_1^{low}$) reduce the number of false positive and false negative results by $\approx 20\%$.

6.1. Correct detection (classification)

Figure 7 presents several examples of correct detection/classification. Correct detection means an image (MRI slice) is classified correctly (malignant or normal) regardless of the location of tumor within the PZ. The segmentation results are divided into three different categories; small malignant region, large malignant region and obscure malignant region. For the first category, the proposed method shows its sensitivity dealing with small malignant regions within the PZ as shown in image 3, 9, 10, 16 and 18. In those images, the proposed method managed to segment malignant regions correctly (in red line) despite their small sizes. In the second category, we show results in detecting and localising malignant region in larger areas. This can be seen in images 1, 4, 7, 13, 14, 15, 17, 19 and 20 where cancers are spread quite substantially within the PZ and some within the CZ. Results in Figure 7 show that these regions were segmented within the expert radiologist's ground truth. Finally (third category), we show results when malignant regions are obscure within the PZ. In image 2, there are three dark regions (left, middle and right) within the PZ and visually it is very difficult to identify which one of those regions is cancerous. As a result, although the proposed method managed to segment the malignant region, there is one false positive region in the middle of the PZ. On the other hand, in image 8 we can visually see that there is no sign of irregularity (the whole PZ looks uniform), which makes the abnormal regions obscured. Other examples of the experimental results can be seen in image 5, 6, 11, 12, 21, 22, 23 and 24.

Figure 8 shows examples of experimental results in normal slices. The PZs in image 25 to 28 show no sign (or small signs) of irregularity which made it easier to identify normal slices.




Figure 7. Malignant slices: prostate capsules are delineated in yellow and central zones and tumors are in green and red, respectively. The detected regions are indicated as the highlighted regions.




Figure 8. Normal slices: prostate capsules are delineated in yellow and central zones in green. The lack of segmented regions on the images indicates no cancer regions present.

6.2. Correct detection with incorrect localisation

This section presents two examples of results where overall classification is correct but the tumor location is incorrect. For localisation, we compare the position of the segmented region based on $(G_1^{low} \cap F_1^{low} \cap F_2^{high} \cap M_1^{low}) \subseteq M_r$, where M_r is a cancerous region within the PZ. In our evaluation, incorrect localisation is when the area of the segmented region is $< 50\%$ within the cancerous region delineated by an expert radiologist. On the other hand, correct localisation means $\geq 50\%$ of the area of the segmented region is within the annotated malignant region. Our method produced 81% (109 slices true positives) correct localisation with respect to the number of malignant

slices (135 samples) which means 6% (8 slices) of malignant cases were classified correctly but tumors were localised incorrectly and the other 18 slices are false negatives. This may have been caused when normal regions have dark or very similar appearance with cancerous regions (low intensity) in the PZ. According to [58] low signal intensity may be seen in the PZ on T2-weighted when blood products may persist after prostate biopsy. Moreover, when the location of the tumor is outside our PZ model (area under the green line in Figure 2). Figure 9 shows examples from our experimental results for correct classification but incorrect localisation.




Figure 9. In both slices (image 29 and 30) the segmentation results show correct overall classification (true positive) but incorrect localisation in comparison to the location of the ground truth.

6.3. False positives and negatives

Figure 10 shows four examples of false positive results from four different prostates. In images 31, 32, 33 and 34, there are clearly dark regions (higher probability of cancer) within the PZ which leads to false positive results. Based on the results in Figure 10, we can visually see that irregularity can occur in some normal slices which makes it hard to differentiate between malignant and normal regions. On the other hand, Figure 11 shows examples of false negative results from four different prostates. The malignant regions show obscure irregularity which lead to false negatives.




Figure 10. False positive results from four different prostates.




Figure 11. False negative results from four different prostates. Tumor regions are delineated in red.

6.4. Parameters justification

Figure 12 shows the justification of our selected parameters. Based on the varying $\sigma = 2, 3, 5, 7$ and 9 , the following kernel sizes $7 \times 7, 9 \times 9, 15 \times 15, 21 \times 21$ and 27×27 are applied, respectively. The results show that better sensitivity (above 80%) is achieved when $2 \leq \sigma \leq 5$. The sensitivity of the proposed method decreases when $\sigma > 5$ due to the level of smoothing applied to images. For instance, higher value of σ would affect (e.g. over-smoothed) the appearance of small malignant regions, hence decreases the sensitivity. On the other hand, the method achieved its highest specificity when $\sigma = 5$. Using $\sigma \leq 5$ still gives similar sensitivity to the other methods in the literature (see Table I). Although varying the σ and kernel size did not change the results significantly, we have shown the quantitative results for the justification of our selected parameters.




Figure 12. Sensitivity, specificity and accuracy using different values of σ and k_s .




Figure 13. Coefficient $\epsilon = 0.75$ produce balanced results, with higher specificity for larger ϵ values and higher sensitivity for lower ϵ values.

Figure 13 shows results using different ϵ values. Our experimental results show that $\epsilon = 0.75$ produced balanced results in terms of accuracy, sensitivity and specificity. A larger ϵ value (e.g. 0.95) reduces the area under the curve (green line in Figure 2), hence most cancerous tissues were missed which increased specificity but reduced the algorithm's sensitivity. On the other hand, a smaller ϵ value (e.g. 0.55) increases the area under the curve (green line in Figure 2). This increased the algorithm's sensitivity (and false positives) because the larger area leads to a higher chance of cancerous tissues being detected.

7. DISCUSSION

Various methods using different frameworks, modalities and features have been proposed in the literature and our method achieved similar results. Nevertheless, it is extremely difficult to make a quantitative comparison due to:

1. Differences in datasets (different modalities such as T2-weighted (T2-W) MRI, diffusion-weighted (DWI) MRI, dynamic contrast enhanced (DCE) MRI, Magnetic resonance spectroscopy (MRS), etc) and frameworks used in the other studies.

2. Absence of public datasets also makes a quantitative comparison of methodologies in the literature difficult. Each team of researchers has their own datasets which cause huge range of variability in terms of noise and image quality.
3. Studies were conducted within different regions of the prostate. For example, some studies were conducted within the prostate PZ only and some took the whole prostate gland into account.
4. Some evaluation was at volume, slice, regions or voxel level.

However, to have an overall qualitative estimate (therefore our comparisons are subjective due to the differences stated above) of the functioning of our method we compared with some of the previous studies in Table I, where methods are categorised as CAD and non-CAD. Studies classified as CAD are methods which are similar to our method (abnormality/malignancy is automatically determined by computer algorithm). On the other hand, non-CAD are studies which involved human readers which, means abnormality/malignancy determination is performed by radiologists/observers.

Table I. Results are ordered based on performance (ACC and AUC are accuracy and area under the curve, respectively), sensitivity and specificity, respectively (all measured in %). For region, each study is either conducted for the whole prostate (WP) or only for the PZ.

Authors	Data size	Performance	Sen	Spe	Category	Modalities	Region
Sung et al. [37]	42	ACC=89	89	89	CAD	DCE	WP
Niaf et al. [49]	30	AUC=89	-	-	CAD	T2-W+DCE+DWI	PZ
Vos at al. [6]	34	ACC=89	-	-	CAD	T2-W+DCE	PZ
Sham et al. [67]	24	<i>F</i> -measure=89	-	-	CAD	T2-W+ADC+DCE	PZ
Ampeliotis et al. [38]	10	ACC=87	-	-	CAD	T2-W+DCE	WP
Our method	37	ACC=86	87	86	CAD	T2-W	PZ
Tiwari et al. [39]	19	ACC=84	-	-	CAD	T2-W+MRS	WP
Chan et al. [70]	15	AUC=84	-	-	CAD	T2-W+ADC+PD+T2 map	PZ
Vos et al. [68]	34	ACC=83	-	-	CAD	DCE	PZ
Rampun et al. [36]	25	ACC=82	81	84	CAD	T2-W	PZ
Artan and Yetik.[10]	15	ACC=82	76	86	CAD	DCE	PZ
Tabesh at al.[47]	29	ACC=81	-	-	CAD	Histological images	WP
Artan et al. [69]	21	ACC=78	74	82	CAD	T2-W+ADC+DCE	PZ
Puech et al. [71]	100	AUC=77	100	45	CAD	DCE	PZ
Kim et al. [55]	20	ACC=75	73	77	Non-CAD	DCE	WP
Han et al. [46]	46	-	96	92	CAD	Ultrasound	WP
Engelbrecht at al.[44]	36	-	93	-	Non-CAD	T2-W+DCE	PZ
Shimofusa at al.[45]	60	-	93	-	Non-CAD	T2-W+DWI	WP
Ito et al. [56]	111	-	87	74	Non-CAD	DCE	WP
Litjens et al. [11]	188	-	83	-	CAD	DWI+DCE	WP
Futterer et al. [40]	34	-	83	83	Non-CAD	T2-W	WP
Reinsberg et al. [41]	42	-	81-93	64-73	Non-CAD	DWI+MRS	WP
Girouin et al. [42]	46	-	78-81	32-56	Non-CAD	DCE	WP
Ocak et al. [18]	50	-	73	88	Non-CAD	DCE	WP
Llobet et al. [43]	303	-	57	61	CAD	Ultrasound	WP
Schlemmer et al. [54]	28	-	-	68	Non-CAD	DCE	WP

Table I presents the experimental results of 26 different methods/studies (including our method) and their accuracies, sensitivities, specificity and modalities (some authors did not report one or two

of these). Although results may be more or less similar, due to several factors mentioned previously the comparisons in this section needs to be treated with caution. Note that every method or study used a different number of patients, modalities and frameworks. However Table I gives a sense of how well the proposed method is capable of analysing cancerous regions in MRI.

In terms of performance, the methods proposed in [6, 37, 49, 67], achieved the highest result of 89%, all used support vector machines (SVM) as classifier. Sung et al. [37] employed SVM as a classifier to distinguish prostate cancer from non-cancerous tissues based on a set of perfusion parameters. Vos et al. [6] used first order statistics of the scalar values of volume as statistical features. Sham et al. [67] developed a decision support system (DSS) which used SVM to generate cancer probability maps from multiparametric MR images and Niaf et al. [49] reported that SVM produced the best result based on a comparison study of four different supervised learning methods (SVM, linear discriminant analysis, k -nearest neighbours and naive Bayes classifiers) based on a feature set derived from grey-level images such as first-order statistics, Haralick features, gradient features, etc. This was followed by a study in [38] which is 2% below the best performance (and 1% above the result of our proposed method) in Table I. Ampeliotis et al. [38] used probabilistic Neural Networks (PNNs) to classify a set of feature vectors extracted from T2- morphological images and T1-W DCE. Other methods [10, 36, 39] achieved more than 80% accuracy. Artan and Yetik [10] developed a random walker (RW) algorithm with automated seed initialisation to segment cancerous region within the PZ using weighted image features and Rampun et al. [36] developed an algorithm which compared local peaks information between the right and left region of prostate's PZ by measuring percentage similarity and Ochiai coefficients to determine the presence of abnormality. A method proposed by Tiwari et al. [39] integrated a semi-supervised multi-kernel (SeSMiK) scheme with a graph embedding framework for data fusion and dimensionality reduction for not only prostate cancer detection but also grading. All non-CAD methods did not report their accuracies except Kim et al. [55] which is, with 75%, below the CAD's average accuracy.

For sensitivity results, Puech et al. [71] reported their CAD software known as 'ProCAD' achieved 100% but only at 45% specificity. The developed software allows for the 2D and multislice 2D contouring of suspicious regions based on a seeded region growing algorithm using standard visualization features such as wash-in and wash-out slopes. Han et al. [46] proposed a method which used a combination of image features and clinical features (e.g. location and shape) and performed SVM to classify cancer and non-cancer region, achieved the second best result of 96% in a single modality. Non-CAD methods [44, 45] used perfusion parameters in multimodalities achieved 93% sensitivity the same as the maximum achieved in [41]. Our method achieved comparable result with Sung et al. [37] (89%) and Ito et al. [56] (87%) in a single modality, while Llobet et al. [43] whose method employed a Hidden Markov Model and k -nn classifiers based on texture descriptors extracted from spatial grey-level dependence metrics and grey-level maps achieved the lowest sensitivity (57%). This may have been influenced by evaluation on a large dataset (4944 ultrasound images). Methods in [41] and [42] show variations of 81-93% and 78-81%, respectively. On the other hand, Niaf et al. [49], who used a combination of three different modalities, achieved 82% similar to the method proposed in [11] based on two modalities. Litjens et al. [11] used a SVM classifier to generate a 3D likelihood map which was used to find points of interest using a local maxima detector. Subsequently, a region is segmented around each local maxima and for every region statistics of the voxel features were calculated and used to discriminate malignant and benign regions.

Finally, in terms of specificity the method of Han et al. [46] obtained the highest result of 92% followed by [37] (89%). Other CAD methods which achieved more than 80% are Artan and Yetik.[10], Rampun et al.[36] and our reported method. The result reported in [18], which used perfusion parameters, achieved 88% is the highest among Non-CAD methods followed by Futterer et al. [40] (83%), while the method in [42] achieved the lowest specificity varying from 32 to 56%. Nevertheless, these comparisons are subjective as accuracy, sensitivity and specificity are highly influenced by several factors mentioned previously. For example although the method proposed in [43] produced the lowest sensitivity, the evaluation is based on 303 prostates. On the other hand, the method proposed by [38] shows higher accuracy on 10 different cases but has not been tested

on larger datasets. Similarly, although Han et al. [46] achieved the highest sensitivity and specificity (based on Table I) but was evaluated on 46 ultrasound images (46 patients).

The challenges of developing a CAD system remain open due to its complexity and limitations both in single and multimodalities imaging. In this study, we are aware that many researchers have attempted to improve results in detecting prostate cancer in MR imaging. Engelbrecht et al. [44] and Sung et al. [37] showed the advantages of using perfusion parameters (e.g. wash in and wash out rates) in detecting prostate cancer in DCE. The method proposed in [6] used a multiparametric MR of T1- and T2-weighted imaging showing better results using a single modality. Shimofusa et al. [45] showed a significant improvement in prostate cancer detection using diffusion-weighted imaging in addition to T2-weighted MRI and reported that sensitivity increased from 87% to 93%. In another study, Reinsberg et al. [41] combined the use of diffusion-weighted MRI and 1H MR Spectroscopy which lead to 93% sensitivity with 73% specificity. The method proposed in [36] used local peak information to detect prostate abnormality by measuring the information difference between left and right PZ regions in a single modality of T2-weighted imaging. Han et al. [46] used clinical knowledge to discriminate the cancer region by location and shape of the region in addition of image features to increase specificity. Moreover, Kim et al. [55] made a comparison between T2-weighted and DCE imaging and reported that DCE has better accuracy (75%) and sensitivity (73%) but 11% lower in specificity (88% in T2-W). Another comparison study was made by Ito et al. [56] between power Doppler ultrasound (PDUS) and DCE imaging and concluded that DCE has higher sensitivity (87%) and specificity (74%) in comparison to PDUS which only achieved 69% and 61%, respectively within the PZ. Llobet et al. [43] who evaluated their method based on 4944 ultrasound images showed similar results with the ones reported on fewer samples by Schlemmer et al. [54]. However, in contrast to the earlier methods, our method is different in the sense that:

1. The proposed method does not need a training phase to be able to discriminate malignant and benign tissues in contrast to the methods in [6, 11, 38, 39, 49, 54, 10].
2. We only used a single modality for abnormality detection which is T2-Weighted MRI. The methods in [44] used multimodality such as diffusion MRI and MR Spectroscopy. Similarly, the method proposed in [6] used a multiparametric MR of T1- and T2-weighted imaging. Engelbrecht et al. [45] suggests that various techniques such as dynamic contrast material enhanced MR imaging, diffusion-weighted imaging, and MR spectroscopy have the potential to improve the detection of prostate cancer. On the other hand [41] combined the use of diffusion-weighted MRI and 1H MR Spectroscopy to get better results in discriminating malignant and normal tissues.
3. The method in [46] used additional clinical knowledge (e.g. location and shape of the region) to discriminate cancer regions in addition of image features while our method only used image features to achieve similar results.
4. Our method used a small number of image features to discriminate malignant and benign regions and produced similar results to the state of art in the literature whereas the methods in [46, 47] used more features.
5. The methods in [18, 37, 44, 45] used various perfusion parameters on a single modality while our method is purely based on image features but still managed to achieve similar results.

The proposed method produced similar accuracy, sensitivity and specificity to the state of art in the literature particularly in single modality T2-Weighted MRI. However, due to various factors mentioned a direct comparison is less appropriate. Several studies [6, 41, 44, 45, 49] have suggested that using image fusion (e.g. combining MRI T2-W with DCE) produces better results in detecting prostate cancer. Nevertheless, other researchers have attempted to make a single predictor (T2-W MRI only) by detecting prostate cancer on the basis of comprehensive analysis of various perfusion parameters, such as in [18] achieved 75% sensitivity and 80% specificity (50 patients). However, without the parameters they achieved higher sensitivity of 94% but much lower specificity of 37%. In another study, Miao et al. [53] reported 76% and 70% sensitivity and specificity, respectively (30 patients) while a study conducted in [42] achieved 50-60% sensitivity and 13-21% specificity in 46 patients. Kim et al. [55] reported 55% sensitivity and 88% specificity (20 patients), and Schlemmer

et al. [54] achieved 79% sensitivity in 28 patients. On the other hand, other investigators [46] have studied the prospect of combining clinical knowledge and image features to detect prostate cancer and achieved similar results. Whether using a single modality, image fusion or using clinical features, none of these methods provide superior results. Therefore developing a CAD tool for prostate cancer detection and localisation remains a challenge.

In our study, one obvious drawback of the proposed method is the risk of classifying correctly with incorrect localisation of the tumor, which could be problematic from a clinical point of view. Secondly, in some cases when the prostate's peripheral zone is almost non-existent, the proposed method is more likely to produce false positives. This is due to the intensities being very similar between the central gland and the malignant region [58]. Finally, if the prostate's shape does not conform to the shape of our prostate model a smaller area of the PZ will be analysed which may increase the chance of malignant regions being missed. Therefore, in order to accommodate these limitations for future work we are planning to use a multiparametric approach (e.g. T2-W+DWI+DCE) instead of stand alone T2-W MRI. This means more image features can be extracted which could help to distinguish malignant and benign tissues. In addition, we intend to cover the whole prostate gland instead of only the PZ.

8. CONCLUSIONS

The proposed method specifies regions which have the highest probability to be malignant (see results in Figure 7), hence help radiologists to perform targeted biopsies and potentially improve the accuracy of prostate cancer diagnosis [57].

In conclusion, we have presented a novel method of prostate cancer detection and localisation within the PZ and successfully applied it on 37 patients. Gaussian and median filters together with probability image information show promising potential to be effective texture descriptors to identify cancer regions within the peripheral region. Our idea, which is based on regions intersection and flexible size of erosion's structuring element, suggest a good potential to reduce false positive and false negative results in the proposed method.

ACKNOWLEDGEMENT

Andrik Rampun would like to thank for the awards given by Aberystwyth University under the Departmental Overseas Scholarship (DOS) and Doctoral Career Development Scholarships (DCDS). This work was funded in part by the NISCHR Biomedical Research Unit for Advanced Medical Imaging and Visualization.

REFERENCES

1. N. Howlader, A. M. Noone, M. Krapcho, J. Garshell, N. Neyman, S. Altekruse, C. Kosary, M. Yu, J. Ruhl, Z. Tatalovich, A. Cho, H. and Mariotto, D. Lewis, H. Chen, E. Feuer, and K. Cronin. (2013) Seer Cancer Statistics Review, 1975-2010, National Cancer Institute. <http://seer.cancer.gov/csr/19752010/>. Accessed 16-October-2013.
2. R. Chou, J. M. Croswell, T. Dana, C. Bougatsos, I. Blazina, R. Fu, K. Gleitsmann, H. C. Koenig, C. Lam, A. Maltz, J. B. Ruge, K. Lin. A Review of the Evidence for the U.S. Preventive Services Task Force, [October 2010] <http://www.uspreventiveservicestaskforce.org/uspstf12/prostate/prostateart.htm> accessed 15th November 2011 11.21 am.
3. <http://www.cancerresearchuk.org/cancer-info/spotcancerearly/>, accessed 7th December 2014 11.35am.
4. K. K. Yu and, H. Hricak. Imaging prostate cancer, *Radiologic Clinics of North America*, vol. 38, pp. 5985, 2000.
5. F. H. Schroder, J. Hugosson, M. J. Roobol, T. L. Tammela, S. Ciatto, V. Nelen, S. Ciatto, V. Nelen, M. Kwiatkowski, M. Lujan, H. Lilja, M. Zappa, L. J. Denis, F. Recker, A. Berenguer, L. Maattanen, C. H. Bangma, G. Aus, A. Villers, X. Rebillard, T. van der Kwast, B. G. Blijenberg, S. M. Moss, H. J. de Koning, and A. Auvinen. for the ERSPC Investigators. Screening and prostate-cancer mortality in a randomized European study. *The New England Journal of Medicine*, vol. 360, pp.1320–1328, 2009.

6. P. C. Vos, T. Hambrock, J. Barentsz, and H. Huisman. Computer-assisted analysis of peripheral zone prostate lesions using T2-weighted and dynamic contrast enhanced T1-weighted MRI, *Physics in Medicine and Biology*, vol. 55, pp.1719–1734, 2010.
7. K. A. Roehl, J. A. Antenor, W. J. Catalona. Serial biopsy results in prostate cancer screening study. *Journal of Urology*, vol. 167, pp. 2435–2439, 2002.
8. <http://www.prostatecancer-riskcalculator.com/other-diagnostic-tests>, accessed 8th December 2014 4.35pm.
9. R. Palmerola, P. Smith, V. Elliot, C. T. Reese, F. B. Mahon, L. E. Harpster, N. Icitovic, J. D. Raman. The digital rectal examination (DRE) remains important outcomes from a contemporary cohort of men undergoing an initial 12-18 core prostate needle biopsy. *Journal of Urology*, vol. 19(6), pp. 6542–6547, 2012.
10. Y. Artan and I. S. Yetik. Prostate cancer localization using multiparametric MRI based on semi-supervised techniques with automated seed initialization, *IEEE Transactions on Information Technology in Biomedicine*, vol. 16(6), pp.1313–1323, 2012.
11. G. J. S. Litjens, P. C. Vos, J. O. Barentsz, N. Karssemeijer, and H. J. Huisman. Automatic computer aided detection of abnormalities in Multi-Parametric prostate MRI. in Proc.SPIE 7963, Medical Imaging 2011: Computer-Aided Diagnosis, 2011.
12. J. J. Fenton, L. Abraham, S.H Taplin, B.M Geller, P. A. Carney, C. D’Orsi, J. G. Elmore, W. E. Barlow; for the Breast Cancer Surveillance Consortium. Effectiveness of computer-aided detection in community mammography practice. *Journal of the National Cancer Institute*, vol. 103, pp. 1152–1161, 2011.
13. I. Sluimer, A. Schilham, M. Prokop, B. Ginneken. Computer analysis of computed tomography scans of the lung: a survey. *IEEE Transactions on Medical Imaging*, vol. 25(385), p. 405, 2006.
14. H. Yoshida, A. H. Dachman. Computer-aided diagnosis for CT colonography. *Semin Ultrasound, CT, MRI*, vol. 25, pp. 419–431, 2004.
15. M. Grana, M. Termenon, A. Savio, A. Gonzalez-Pinto, J. Echeveste, J. M. Prez, A. Besga. Computer Aided Diagnosis system for Alzheimer Disease using brain Diffusion Tensor Imaging features selected by Pearsons correlation. *Neuroscience Letters*, vol. 502, pp. 225–229, 2011, doi:10.1016/j.neulet.2011.07.049.
16. S. B. Edge, D. R. Byrd, C. Compton, A. G. Fritz, F. L. Greene, and A. Trotti (2010). *AJCC Cancer Staging Manual* (7th Edition). Springer., Chicago, US.
17. H. Ito, K. Kamoi, K. Yokoyama, K. Yamada, and T. Nishimura. Visualization of prostate cancer using dynamic contrast-enhanced mri: comparison with transrectal power doppler ultrasound, *British Journal of Radiology*, vol. 76 (909), pp. 617–624, 2003.
18. I. Ocak, M. Bernardo, G. Metzger, T. Barrett, P. Pinto, P. S. Albert, and P. L. Choyke. Dynamic contrast-enhanced mri of prostate cancer at 3 t: a study of pharmacokinetic parameters, *American Journal of Roentgenology*, vol. 189 (4), pp.W192–W201, 2007.
19. A. A. T. Bui and R. K. Taira (2010). *Medical Imaging Informatics* (pp. 215). Springer., New York, US.
20. D. Garcia. Robust smoothing of gridded data in one and higher dimensions with missing values. *Computational Statistics and Data Analysis*, vol. 54(4), pp. 1167–1178, 2010.
21. K. Cabeen and P. Gent (2012). Image compression and the Discrete Cosine Transform. <http://www.lokminglui.com/dct.pdf>, accessed 11-August-2013.
22. A. B. Watson (1993). Dctune: A technique for visual optimization of dct quantization matrices for individual images. *Society for Information Display Digest of Technical Papers*, XXIV:946–949.
23. P. Getreuer. A Survey of Gaussian Convolution Algorithms, *Image Processing On Line*, 2012. DOI: 10.5201/ipol.2013.87.
24. M. A. Haidekker (2010). *Advanced Biomedical Image Analysis* (pp. 32). John Wiley & Son., New Jersey, US.
25. R. Bourne (2010). *Fundamentals of Digital Imaging in Medicine*(pp. 155). Springer., London, UK.
26. R. C. Gonzalez and R. E. Woods (2007). *Digital Image Processing* (3rd Edition). Practice Hall., New Jersey, US.
27. J. L. Semmlow (2004). *Biosignal and Medical Image Processing* (pp. 312). Marcel Dekker Inc., New York, US.
28. R. Lasaponara and N. Masini (2012). *Satellite Remote Sensing: A New Tool for Archeology* (pp. 51). Springer, New York, US.
29. D. Q. Zhang and S. C. Chen. A novel kernelized fuzzy cc-means algorithm with application in medical image segmentation, *Artificial Intelligence in Medicine*, vol. 32(1), pp. 37–50, 2014.
30. <http://homepages.inf.ed.ac.uk/rbf/HIPR2/median.htm>, accessed 23rd January 2014.
31. E. J. Halpern, D. L. Cochlin, and B. Goldberg, *Imaging of the Prostate*. London, UK: Martin Dunitz Ltd., 2002.
32. M. B. Garnick, A. MacDonald, R. Glass, and S. Leighton, *Harvard Medical School 2012 Annual Report on Prostate Diseases*. Harvard, US: Harvard Medical School, 2012.
33. D. T. Ginat, S. V. Destounis, R. G. Barr, B. Castaneda, J. G. Strang, and D. J. Rubens. Us elastography of breast and prostate lesions, *Radiographics*, vol. 29 (7), pp. 2007–2016, 2009.
34. H. Choi, E. Loyer, H. Kaur and P. M. Silverman. Imaging Neoplasms of the Abdomen and Pelvis. In: Kufe DW, Pollock RE, Weichselbaum RR, et al., editors. *Holland-Frei Cancer Medicine*. 6th edition. Hamilton (ON): BC Decker; 2003. Chapter 36d. Available from: <http://www.ncbi.nlm.nih.gov/books/NBK13197/>
35. S. S. Taneja. Imaging in the diagnosis and management of prostate cancer, *Reviews in Urology*, vol. 6 (3), p. 101, 2004.
36. A. Rampun, and R. Zwigelaar. Detection of Prostate Abnormality Within the Peripheral Zone Using Local Peak Information. In 3rd International Conference on Pattern Recognition Applications and Methods (ICPRAM 14), 2014.
37. Y. S. Sung, H.-J. Kwon, B. W. Park, G. Cho, C. K. Lee, K.-S. Cho, and J. K. Kim. Prostate cancer detection on dynamic contrast-enhanced mri: Computer-aided diagnosis versus single perfusion parameter maps, *American Journal of Roentgenology*, vol. 197 (5), pp. 1122–1129, 2011.
38. D. Ampeliotis, A. Antonakoudi, K. Berberidis, and E. Z. Psarakis. Computer aided detection of prostate cancer using fused information from dynamic contrast enhanced and morphological magnetic resonance images, in *IEEE International Conference on Signal Processing and Communications(ICSPC 2007)*, 2007.

39. P. Tiwari, J. Kurhanewicz, M. Rosen, and A. Madabhushi. Semi supervised multi kernel (sesmik) graph embedding: identifying aggressive prostate cancer via magnetic resonance imaging and spectroscopy, in *Medical Image Computing and Computer-Assisted Intervention MICCAI* 2010. Springer, 2010, pp. 666–673.
40. J. J. Futterer, S. W. T. P. J. Heijmink, T. W. J. Scheenen, J. Veltman, H. J. Huisman, P. Vos, C. A. H. de Kaa, J. A. Witjes, P. F. M. Krabbe, A. Heerschap, and J. O. Barentsz. Prostate cancer localization with dynamic contrast-enhanced mr imaging and proton mr spectroscopic imaging, *Radiology*, vol. 241 (2), pp. 449–458, 2006, pMID: 16966484, [Online]. Available: <http://pubs.rsna.org/doi/abs/10.1148/radiol.2412051866>.
41. S. A. Reinsberg, G. S. PaVosyne, S. F. Riches, S. Ashley, J. M. Brewster, V. A. Morgan, and N. M. deSouza. Combined use of Diffusion-Weighted MRI and 1H MR Spectroscopy to increase accuracy in prostate cancer detection, *American Journal of Roentgenology*, vol. 188 (1), pp. 1122–1129, 2007.
42. N. Girouin, F. Mege-Lechevallier, A. T. Senes, A. Bissery, M. Rabilloud, J. Marechal, M. Colombel, D. Lyonnet, and O. Rouviere. Prostate dynamic contrast-enhanced mri with simple visual diagnostic criteria: is it reasonable?, *European Radiology*, vol. 17 (6), pp. 1498–1509, 2007.
43. R. Llobet, C. Juan, P. Cortes, A. Juan, and A. Toselli. Computer-aided detection of prostate cancer, *International Journal of Medical Informatics*, vol. 76(7), pp. 547–556, 2007.
44. M. R. Engelbrecht, H. J. Huisman, R. J. Laheij, G. J. Jager, G. J. van Leenders, C. A. Hulsbergen-Van De Kaa, J. J. de la Rosette, J. G. Blickman, J. O. Barentsz. Discrimination of prostate cancer from normal peripheral zone and central gland tissue by using dynamic contrast-enhanced MR imaging, *Radiology*, vol. 229, pp. 248–254, 2003.
45. R. Shimofusa, H. Fujimoto, H. Akamata, K. Motoori, S. Yamamoto, T. Ueda, H. Ito. Diffusion-weighted imaging of prostate cancer, *Journal of Computer Assisted Tomography*, vol. 29, pp. 149–153, 2005.
46. S. Han, H. Lee and J. Choi. Computer-aided prostate cancer detection using texture features and clinical features in ultrasound image, *Journal of Digital Imaging*, vol. 21, no. Suppl 1, pp. 121–133, 2008.
47. A. Tabesh, M. Teverovskiy, H. Y. Pang, V. P. Kumar, D. Verbel, A. Kotsianti and O. Saidi. Multifeature prostate cancer diagnosis and Gleason grading of histological images, *IEEE Transactions on Medical Imaging*, vol. 26, no. 10, pp. 1366–1378, 2007.
48. K. Nguyen, A. K. Jain, and B. Sabata. Prostate cancer detection: Fusion of cytological and textural features, *Journal of Pathology Informatics*, vol. 2(3), 2011.
49. E. Niaf, O. Rouviere, F. Mege-Lechevallier, F. Bratan and C. Lartzien. Computer-aided diagnosis of prostate cancer in the peripheral zone using multiparametric MRI, *Physics in Medicine and Biology*, vol. 57, pp. 3833–3851, 2012.
50. P. C. Vos, T. Hambroek, C. A. Hulsbergen-van de Kaa, J. J. Futterer, J. O. Barentsz, H. J. Huisman. Computerized analysis of prostate lesions in the peripheral zone using dynamic contrast enhanced MRI, *Medical Physics*, vol. 35, pp. 888–899, 2008.
51. M. Hachama, A. Desolneux, F. Richard. Combining registration and abnormality detection in mammography. Proceedings of International Workshop on Biomedical Image Registration; Utrecht, The Netherlands. July 9–11, 2006; pp. 178–185.
52. M. Yin, R. Dhir, A. V. Parwani. Diagnostic utility of p501s (prostein) in comparison to prostate specific antigen (PSA) for the detection of metastatic prostatic adenocarcinoma, *Diagnostic Pathology*, vol. 2(41), 2007.
53. H. Miao, H. Fukatsu, T. Ishigaki. Prostate cancer detection with 3-T MRI: comparison of diffusion weighted and T2-weighted imaging, *European Journal of Radiology*, vol. 61, pp. 297–302, 2007.
54. H. P. Schlemmer, J. Merkle, R. Grobholz. Can preoperative contrast-enhanced dynamic MR imaging for prostate cancer predict microvessel density in prostatectomy specimens? *European Radiology*, vol. 14, pp. 309–317, 2004.
55. K. C. Kim, B. K. Park and B. Kim. Localization of prostate cancer using 3T MRI: comparison of T2-weighted and dynamic contrast-enhanced imaging, *Journal of Computer Assisted Tomography*, vol. 30, pp. 7–11, 2006.
56. H. Ito, K. Kamoi, K. Yokoyama, K. Yamada and T. Nishimura. Visualization of prostate cancer using dynamic contrast-enhanced MRI: comparison with transrectal power Doppler ultrasound, *The British Journal of Radiology*, vol. 76, pp. 617–624, 2007.
57. <http://www.bbc.co.uk/news/health-26970132>, accessed 4th April 2014.
58. Y. J. Choi, J. K. Kim, N. Kim, K. W. Kim, E. K. Choi and K. S. Cho. Functional MR imaging of prostate cancer, *Radiographics*, vol. 27(1), pp. 63–75, 2007.
59. R. E. Hendrick (2007). *Breast MRI: Fundamentals and Technical Aspects*. Springer, Chicago, US.
60. A. M. Wink and J. B. T. M. Roerdink. Denoising functional MR images: A comparison of wavelet denoising and Gaussian smoothing, *IEEE Transactions on Medical Imaging*, vol. 23 (3), pp. 374–387, 2004.
61. K. Xiao, S. H. Ho and Q. A. Salih. A study: segmentation of lateral ventricles in brain MRI using fuzzy c-means clustering with Gaussian smoothing. In Proceedings of the Joint Rough Set Symposium, Lecture Notes in Computer Science, vol. 4482, pp. 161–170, 2007.
62. R. Zwiggelaar, C. J. Taylor and C. M. E. Rubin (1999). Detection of the central mass of spiculated lesions - signature normalisation and model data aspects, *Information Processing in Medical Imaging*, Springer-Verlag, London, UK, pp. 406–411.
63. S. Gharge and H. B. Kekre (2013). *Segmentation of Medical Images*. Ph.D. Thesis. Mukesh Patel School of Technology Management and Engineering: India.
64. M. Bydder, D. J. Larkman, J. V. Hajnal. Combination of signals from array coils using image-based estimation of coil sensitivity profiles, *Magnetic Resonance in Medicine*, vol. 47, pp. 539–548, 2002.
65. J. C. Russ (2002). *The Image Processing Handbook* (4th Edition). CRC Press, US.
66. Z. Chen and R. Zwiggelaar. A Modified Fuzzy C-Means Algorithm for Breast Tissue Density Segmentation in Mammograms. In Proceedings of the 10th International Conference on Information Technology and Applications in Biomedicine, 3–5 November, Corfu, 2010.
67. V. Shah, B. Turkbey, H. Mani, Y. Pang, T. Pohida, M. J. Merino, P. A. Pinto, P. L. Choyke and M. Bernardo. Decision support system for localizing prostate cancer based on multiparametric magnetic resonance imaging, *Medical Physics*, vol. 39 (7), pp. 4093–4103, 2012.
68. P. C. Vos, T. Hambroek, C. A. H. van de Kaa, J. J. Futterer, J. O. Barentsz, and H. J. Huisman. Computerized analysis of prostate lesions in the peripheral zone using dynamic contrast enhanced MRI, *Medical Physics*, vol. 35 (3), pp.

- 888–899, 2008.
69. Y. Artan, M. A. Haider, D. L. Langer, T. H. van der Kwast, A. J. Evans, Y. Yang, M. N. Wernick, J. Trachtenberg and I. S. Yetik. Prostate cancer localization with multispectral MRI using cost-sensitive support vector machines and conditional random fields, *IEEE Trans Image Process.* vol. 19(9), pp. 2444–2455, 2010.
 70. I. Chan, W. Wells, R. V. Mulkern, S. Haker, J. Zhang, K. H. Zou, S. E. Maier and C. M. Tempny. Detection of prostate cancer by integration of line-scan diffusion, T2-mapping and T2-weighted magnetic resonance imaging; a multichannel statistical classifier, *Medical Physics*, vol. 30 (9), pp. 2390–2398, 2003.
 71. P. Puech, N. Betrouni, N. Makni, A.-S. Dewalle, A. Villers, and L. Lemaitre. Computer-assisted diagnosis of prostate cancer using DCE-MRI data: design, implementation and preliminary results, *International Journal of Computer Assisted Radiology and Surgery*, vol. 4 (1), pp. 1–10, 2009.
 72. Y. J. Choi, J. K. Kim, N. Kim, K. W. Kim, E. K. Choi, and K.-S. Cho. Functional MR Imaging of Prostate Cancer, *RadioGraphics*, vol. 27 (1), pp. 63–75, 2007.
 73. N. Makni, A. Iancu, O. Colot, P. Puech, S. Mordon and N. Betrouni. Zonal segmentation of prostate using multispectral magnetic resonance images, *Medical Physics*, vol. 38, pp. 6093–6105, 2011.
 74. J. Carlsson, G. Helenius, M. G. Karlsson, O. Andren, K. Klinga-Levan and B. Olsson. Differences in microRNA expression during tumor development in the transition and peripheral zones of the prostate. *BMC Cancer*, vol. 13 (1), pp. 6093–6105, 2013.
 75. D. L. Pham, C. Xu and J. L. Prince. Current Methods in Medical Image Segmentation. *Annual Review of Biomedical Engineering*. vol. 2. pp. 315–338, 2000.
 76. D. Doi. Computer-Aided Diagnosis in Medical Imaging: Historical Review, Current Status and Future Potential. *Computerized Medical Imaging and Graphics*. vol. 31 (4-5). pp. 198–211, 2007.
 77. R. B. Dubey, M. Hanmandlu and S. K. Gupta. Region Growing for MRI Brain Tumor Volume Analysis. *Indian Journal of Science and Technology*. vol. 2(9), 2009.
 78. A. Tsai, A. Yezzi Jr., W. Wells, C. Tempny, D. Tucker, A. Fan, W. E. Grimson and A. Willsky. A Shape-Based Approach to the Segmentation of Medical Imagery Using Level Sets. *IEEE Trans on Medical Imaging* vol. 22(2). pp. 103–154, 2003.
 79. J. Liu, J. Chen, X. Liu, L. Chun and J. Tang. Mass Segmentation in Mammograms Based on Improved Level Set and Watershed Algorithm. *Advanced Intelligent Computing Theories and Applications. With Aspects of Artificial Intelligence Lecture Notes in Computer Science*, vol. 6839, 2012, pp 502–508.
 80. J. Shi, B. Sahiner, H. P. Chan, J. Ge, L. Hadjiiski, M. A. Helvie, A. Nees, Y. T. Wu, J. Wei, C. Zhou, Y. Zhang and J. Cui. Characterization of mammographic masses based on level set segmentation with new image features and patient information, *Medical Physics*, vol. 35(1), pp. 280–290, 2008.
 81. S. Y. Yeo, X. Xie, I. Sazonov and P. Nithiarasu. Segmentation of biomedical images using active contour model with robust image feature and shape prior, *International Journal for Numerical Methods in Biomedical Engineering*, vol. 30(2), pp. 232–48, 2014.
 82. J. Sachdeva, V. Kumar, I. Gupta, N. Khandelwal and C. K. Ahuja. A novel content-based active contour model for brain tumor segmentation, *Magnetic Resonance Imaging*, vol. 30(5), pp. 694–715, 2012.
 83. B. Tanoori, Z. Azimifar, A. Shakibafar and S. Katebi. Brain volumetry: an active contour model-based segmentation followed by SVM-based classification, *Computers in Biology and Medicine*, vol. 41(8), pp. 619–632, 2011.
 84. V. Pedoia and E. Binaghi. Automatic MRI 2D brain segmentation using graph searching technique, *Magnetic Resonance Imaging*, vol. 29(9), pp. 887–904, 2003.
 85. K. Li, X. Wu, D. Z. Chen and M. Sonka. Optimal Surface Segmentation in Volumetric Images A Graph-Theoretic Approach, *IEEE Pattern Analysis and Machine Intelligence*, vol. 29(9), pp. 119–134, 2006.
 86. N. Fordyce, S. McKenna, C. Hacker and J. Lucocq. Finding Golgi stacks in Electron Micrographs. In *Medical Image Understanding and Analysis (MIUA 14)*, 2014.
 87. Y. Wang, G. W. Wei and S. Yang. Partial differential equation transform - Variational formulation and Fourier analysis, *International Journal for Numerical Methods in Biomedical Engineering*, vol. 27(12), pp. 1996–2020, 2011.
 88. G. Gerig, O. Kubler, R. Kikinis, F. A. Jolesz. Nonlinear anisotropic filtering of MRI data, *IEEE Transactions on Medical Imaging*, vol. 11 (2), pp. 221–232, 1992.
 89. J. O. Barentsz, J. Richenberg, R. Clements, P. Choyke, S. Verma, G. Villeirs, O. Rouviere, V. Logager, and J. J. Fitterer. European Society of Urogenital Radiology, ESUR prostate MR guidelines 2012. *European Radiology*, vol. 22, pp. 746–757, 2012.
 90. S. E. Viswanath, N. B. Bloch, J. C. Chappelow, R. Toth, N. M. Rofsky, E. M. Genega, R. E. Lenkinski, and A. Madabhushi. Central gland and peripheral zone prostate tumors have significantly different quantitative imaging signatures on 3 tesla endorectal, in vivo T2-weighted MR imagery, *Journal of Magnetic Resonance Imaging*, vol. 36, pp. 213–224, 2012.
 91. X. Liu, M. A. Haider and S. Yetik. Automated Prostate Cancer Localization with MRI without the need of manually extracted Peripheral Zone, *Medical Physics*, vol. 38(6), pp. 2986–2994, 2011.
 92. G. Litjens, O. Debats, J. Barentsz, N. Karssemeijer and H. Huisman. Computer Aided Detection of Prostate Cancer in MRI, *IEEE Transactions on Medical Imaging*, vol. 33(5), pp. 1083–1092, 2014.
 93. E. A. El-Dahshana, H. M. Mohsenc, K. Revettd and A. M. Salemf. Computer-aided diagnosis of human brain tumor through MRI: A survey and a new algorithm, *Expert Systems with Applications*, vol. 41(11), pp. 5526–5545, 2014.
 94. Y. Zhu, S. Williams, R. Zwigelaar. A hybrid ASM approach for sparse volumetric data segmentation, *Pattern Recognition and Image Analysis, Lecture Notes in Computer Science*, Vol. 17(2), pp. 252–258, 2007.
 95. J. Jones, X. Xianghua, E. Ehab. Combining region-based and imprecise boundary-based cues for interactive medical image segmentation, *International Journal for Numerical Methods in Biomedical Engineering*, vol. 30(12), pp. 1649–1666, 2014.
 96. J. Fiot, L. D. Cohen, P. Raniga, J. Fripp. Combining region-based and imprecise boundary-based cues for interactive medical image segmentation, *International Journal for Numerical Methods in Biomedical Engineering*, vol. 29(9), pp. 905–915, 2013.